

RADIOLOGICAL CONTAMINATION SURVEY FORM

Facility: Section 12 Mine Release Survey Performed By: Sam Paterniti Date: 07/09/2021
 Survey Description: Hoist House Drum 1 RWP: n/a

Instrument Make/Model:	Ludlum Model 19		Ludlum Model 2360 with Ludlum Model 43-93				Ludlum Model 2929 with Ludlum Model 43-10-1			
Instrument Serial No.:	144026		347133/PR399491				152268/PR156426			
Calibration Due Date:	12/16/2021		5/27/2022				7/1/2022			
Survey Background:	gamma (µR/hr):	30	alpha (cpm):	1	beta (cpm):	457	alpha (cpm):	1	beta (cpm):	66
Calibration Background: ⁽¹⁾	gamma (µR/hr):	NA	alpha (cpm):	10	beta (cpm):	289	alpha (cpm):	4	beta (cpm):	64
Calibration Source: ⁽¹⁾	Cs-137 (µCi)	NA	Th-230 (epm)	1481	Tc-99 (epm)	5002	Th-230 (epm)	2443	Tc-99 (epm)	5259
Total Efficiency: ⁽²⁾	n/a		(cpm/epm)	0.10	(cpm/epm)	0.06	(cpm/epm)	0.16	(cpm/epm)	0.07

#	Description/Location	Exposure Rate		Total Alpha		Total Beta		Removable Alpha		Removable Beta	
		Gross Gamma (µR/hr)	Net Gamma (µR/hr)	Gross (cpm)	Surface Activity (dpm/100 cm ²)	Gross (cpm)	Surface Activity (dpm/100 cm ²)	Gross (cpm)	Activity (dpm/100 cm ²)	Gross (cpm)	Activity (dpm/100 cm ²)
1	Top Surface of Motor	35	5	29	290	438	0	7	38	71	70
2	Wire (Top of Spool)	32	2	12	114	519	988	5	25	71	70
3	Casing (Bottom of Interior)	32	2	14	135	471	223	5	25	61	0
4	Drum Shaft (Interior)	33	3	18	176	449	0	3	13	72	84
5	Wire (Bottom of Spool)	30	0	12	114	582	1991	3	13	60	0
6	Casing (Exterior, Top)	28	0	14	135	345	0	3	13	58	0
7											
8											
9											
10											

Comments: Survey backgrounds taken on site. For source information see function check forms.
Technician Signature/Date: <i>Sam Paterniti</i> 7/11/21
Reviewer Signature/Date: <i>[Signature]</i> 7/12/21




Notes:

- (1) This data is taken from the respective instrument calibration sheet. Instrument efficiency does not change as long as equipment function checks indicate normal operation.
- (2) The instrument total efficiency is calculated per NUREG 1575; Total Efficiency = Instrument Efficiency × Source Efficiency. There is no instrument efficiency calculated for the Ludlum Model 19.

EDF-5 RADIOLOGICAL CONTAMINATION SURVEY FORM (Supplement)

Facility: Drum 1 in Hoist House and Mechanical Shop

Date: 07/09/2021



RADIOLOGICAL CONTAMINATION SURVEY FORM

Facility: **Section 12 Mine**

Release Survey Performed By: **Sam Paterniti**

Date: **07/09/2021**

Survey Description: **Hoist House Drum 2**

RWP:

Instrument Make/Model:	Ludlum Model 19		Ludlum Model 2360 with Ludlum Model 43-93				Ludlum Model 2929 with Ludlum Model 43-10-1			
Instrument Serial No.:	144026		347133/PR399491				152268/PR156426			
Calibration Due Date:	12/16/2021		5/27/2022				7/1/2022			
Survey Background:	gamma (µR/hr):	30	alpha (cpm):	1	beta (cpm):	457	alpha (cpm):	1	beta (cpm):	66
Calibration Background: ⁽¹⁾	gamma (µR/hr):	NA	alpha (cpm):	10	beta (cpm):	289	alpha (cpm):	4	beta (cpm):	64
Calibration Source: ⁽¹⁾	Cs-137 (µCi)	NA	Th-230 (epm)	1481	Tc-99 (epm)	5002	Th-230 (epm)	2443	Tc-99 (epm)	5259
Total Efficiency: ⁽²⁾	n/a		(cpm/epm)	0.10	(cpm/epm)	0.06	(cpm/epm)	0.16	(cpm/epm)	0.07

#	Description/Location	Exposure Rate		Total Alpha		Total Beta		Removable Alpha		Removable Beta	
		Gross Gamma (µR/hr)	Net Gamma (µR/hr)	Gross (cpm)	Surface Activity (dpm/100 cm²)	Gross (cpm)	Surface Activity (dpm/100 cm²)	Gross (cpm)	Activity (dpm/100 cm²)	Gross (cpm)	Activity (dpm/100 cm²)
1	Top Surface of Motor	20	0	18	176	463	96	3	13	65	0
2	Wire (Top of Spool)	24	0	18	176	386	0	7	38	69	42
3	Casing (Bottom of Interior)	24	0	13	124	570	1800	4	19	72	84
4	Drum Shaft (Interior)	24	0	28	279	438	0	6	31	70	56
5	Wire (Bottom of Spool)	22	0	8	72	405	0	2	6	55	0
6	Casing (Exterior, Top)	18	0	3	21	364	0	2	6	53	0
7											
8											
9											
10											



Comments: Survey backgrounds taken on site. For source information see function check forms.				
Technician Signature/Date:	<i>Sam Paterniti</i>		July 11, 2021	
Reviewer Signature/Date:	<i>[Signature]</i>		July 12, 2021	

Notes:
 (1) This data is taken from the respective instrument calibration sheet. Instrument efficiency does not change as long as equipment function checks indicate normal operation.
 (2) The instrument total efficiency is calculated per NUREG 1575; Total Efficiency = Instrument Efficiency × Source Efficiency. There is no instrument efficiency calculated for the Ludlum Model 19.

EDF-5 RADIOLOGICAL CONTAMINATION SURVEY FORM (Supplement)

Facility: **Drum 2 in Hoist House**

Date: **07/09/2021**



RADIOLOGICAL CONTAMINATION SURVEY FORM

Facility: **Section 12 Mine**

Release Survey Performed By: **Sam Paterniti**

Date: **07/09/2021**

Survey Description: **Bull Motor and Gear**

RWP:

Instrument Make/Model:	Ludlum Model 19		Ludlum Model 2360 with Ludlum Model 43-93				Ludlum Model 2929 with Ludlum Model 43-10-1			
Instrument Serial No.:	144026		347133/PR399491				152268/PR156426			
Calibration Due Date:	12/16/2021		5/27/2022				7/1/2022			
Survey Background:	gamma (µR/hr):	30	alpha (cpm):	1	beta (cpm):	457	alpha (cpm):	1	beta (cpm):	66
Calibration Background: ⁽¹⁾	gamma (µR/hr):	NA	alpha (cpm):	10	beta (cpm):	289	alpha (cpm):	4	beta (cpm):	64
Calibration Source: ⁽¹⁾	Cs-137 (µCi)	NA	Th-230 (epm)	1481	Tc-99 (epm)	5002	Th-230 (epm)	2443	Tc-99 (epm)	5259
Total Efficiency: ⁽²⁾	n/a		(cpm/epm)	0.10	(cpm/epm)	0.06	(cpm/epm)	0.16	(cpm/epm)	0.07

#	Description/Location	Exposure Rate		Total Alpha		Total Beta		Removable Alpha		Removable Beta	
		Gross Gamma (µR/hr)	Net Gamma (µR/hr)	Gross (cpm)	Surface Activity (dpm/100 cm ²)	Gross (cpm)	Surface Activity (dpm/100 cm ²)	Gross (cpm)	Activity (dpm/100 cm ²)	Gross (cpm)	Activity (dpm/100 cm ²)
1	Exterior Gear	18	0	7	62	411	0	7	38	74	112
2	Casing (Top)	25	0	3	21	339	0	6	31	70	56
3	Interior Gear	25	0	9	83	582	1991	3	13	73	98
4	Steel Lower Support Arm	25	0	8	72	411	0	5	25	69	42
5	Motor Cap	28	0	6	52	445	0	3	13	65	0
6	Casing (Exterior, bottom)	28	0	11	104	323	0	4	19	89	321
7											
8											
9											
10											

Comments: Survey backgrounds taken on site. For source information see function check forms.				
Technician Signature/Date: <i>Sam Paterniti</i> July 11, 2021				
Reviewer Signature/Date: <i>[Signature]</i> July 12, 2021				



Notes:

- (1) This data is taken from the respective instrument calibration sheet. Instrument efficiency does not change as long as equipment function checks indicate normal operation.
- (2) The instrument total efficiency is calculated per NUREG 1575; Total Efficiency = Instrument Efficiency × Source Efficiency. There is no instrument efficiency calculated for the Ludlum Model 19.

EDF-5 RADIOLOGICAL CONTAMINATION SURVEY FORM (Supplement)

Facility: **Bull Drum in Hoist**


Date: **07/09/2021**



Administrative Building



All measurement are in micro roentgen per hour ($\mu\text{R/hr}$) collected with a Ludlum Model 19.





Administrative Building



Room 1



Room 2



Room 3




Hallway



Room 4

Hoist House and Mechanical Shop

All measurement are in micro roentgen per hour ($\mu\text{R/hr}$) collected with a Ludlum Model 19.





Hoist House and Mechanical Building



Room 1 (Looking South)



Room 1 (Looking North)



Room 2 (Looking North)



Room 3 (Looking South)



Room 3 (Looking North)



Single-Channel Function Check Log

Environmental Restoration Group, Inc.
8809 Washington St. NE, Suite 150
Albuquerque, NM 87113
(505) 298-4224

METER	
Manufacturer:	Ludlum
Model:	Model 19
Serial No.:	144026
Cal. Due Date:	16-Dec-21

DETECTOR	
Manufacturer:	
Model:	
Serial No.:	
Cal. Due Date:	

Comments:
Function checked at ERG/Sam P office.
Scaler Count Time: n/a
Distance To Source: ERG Cal jig, approx 5-in.

Source: Cs-137 Serial No.: 5309-04 Activity: 5.7 uCi
 Emission Rate: 5.7 cpm/emissions Source Date: Jig 16-11-SP

NOTE: For use as needed. Acceptable upper/lower net counts, detector total efficiency, and detector MDA calculated on ERG Form ITC.201.B.

Acceptable Upper Net Counts:	<u>NA</u>	Total Efficiency (E _t):	<u>NA</u>
Acceptable Lower Net Counts:	<u>NA</u>	MDA (dpm/100-cm ²):	<u>NA</u>

Date	Time	Battery	High Voltage	Threshold	Source Counts	BKG Counts	Net Counts	Initials	Note(s)
7/09/21	0715	ok	NA	NA	400	12	388	SP	
7/09/21	1817	ok	NA	NA	400	12	388	SP	
<i>[Large handwritten signature/initials 'SP' with a large arrow pointing across the table]</i>									

Reviewed by: *C. Far*

Review Date: July 12, 2021



Dual-Channel Function Check Log

Environmental Restoration Group, Inc.
8809 Washington St. NE, Suite 150
Albuquerque, NM 87113
(505) 298-4224

METER	
Manufacturer:	2360 Ludlum
Model:	2360
Serial No.:	347133
Cal. Due Date:	27-May-22

DETECTOR	
Manufacturer:	Ludlum
Model:	43-93
Serial No.:	PR399491
Cal. Due Date:	27-May-22

Comments:	
Function checked at ERG/Sam P office.	
Scaler Count Time(s): 1-min.	
Distance To Source: Contact	

Alpha Source: Th-230 Serial No.: 7132-10 Emission Rate: 7550 dpm / 3830 cpm Source Date: 06/15/11
 Beta Source: Tc-99 Serial No.: 93TC4700530 Emission Rate: 29000 dpm / 18100 cpm Source Date: 01/29/93

NOTE: For use as needed. Acceptable upper/lower net counts, detector total efficiency, and detector MDA calculated on ERG Form ITC.202.B.

	α Alpha	β Beta
Total Efficiency (E):	0.10	0.06
MDA (dpm/100-cm ²):		

	Alpha	Beta
Acceptable Upper Net Counts:	NA	NA
Acceptable Lower Net Counts:	NA	NA

Date	Time	Battery	High Voltage	Alpha Counts		Beta Counts		BKG Counts		Net Counts		Initials
				α Alpha	α Beta	β Alpha	β Beta	Alpha	Beta	α Alpha	β Beta	
7/9/21	0720	ok	625	1483	520	8	4930	7	320	1476	4610	SP
7/9/21	1822	ok	625	1434	503	6	5008	5	294	1429	4714	SP

Reviewed by: C. Far

Review Date: July 12, 2021



Dual-Channel Function Check Log

Environmental Restoration Group, Inc.
8809 Washington St. NE, Suite 150
Albuquerque, NM 87113
(505) 298-4224

METER	
Manufacturer:	Ludlum
Model:	2929
Serial No.:	152263
Cal. Due Date:	1-July-22

DETECTOR	
Manufacturer:	Ludlum
Model:	43-10-1
Serial No.:	PR 156426
Cal. Due Date:	1-July-22

Comments:	
Function checked at ERG/Sam P office.	
Scaler Count Time(s):	1-min
Distance To Source:	In-tray

Alpha Source: Th-230 Serial No.: 7132-10 Emission Rate: 7550 dpm/3830 cpm cpm/emissions Source Date: 06/15/11
 Beta Source: Tc-99 Serial No.: 93TC4700530 Emission Rate: 29000 dpm/18,100 cpm/emissions Source Date: 01/29/93

NOTE: For use as needed. Acceptable upper/lower net counts, detector total efficiency, and detector MDA calculated on ERG Form ITC.202.B.

	α Alpha	β Beta
Total Efficiency (E):	0.16	0.07
MDA (dpm/100-cm ²):		

	Alpha	Beta	
Acceptable Upper Net Counts:	—	—	NA
Acceptable Lower Net Counts:	—	—	NA

Date	Time	Battery	High Voltage	Alpha Counts		Beta Counts		BKG Counts		Net Counts		Initials
				α Alpha	α Beta	β Alpha	β Beta	Alpha	Beta	α Alpha	β Beta	
7/9/21	0715	ok	950	2565	230	2	5230	5	61	2563 ^{SP}	5259	SP
7/9/21	1815	ok	950	2504	251	3	5168	2	67	2502	5101	SP

2560

Reviewed by: C. Far

Review Date: July 12, 2021